



Product Service

# CERTIFICATE

No. Z1 022430 0054 Rev. 00

**Holder of Certificate:** **Alu Rehab AS**  
 Bedriftsvegen 23  
 4353 Klepp Stasjon  
 NORWAY

**Certification Mark:**



**Product:** **Standard wheelchairs**  
**Manual wheelchair**

The product was tested on a voluntary basis and complies with the essential requirements. The certification mark shown above can be affixed on the product. It is not permitted to alter the certification mark in any way. In addition the certification holder must not transfer the certificate to third parties. See also notes overleaf.

**Test report no.:** 713096380

**Valid until:** 2024-01-29

**Date,** 2019-01-30

  
 ( Torsten Zimmer )



Product Service

# CERTIFICATE

No. Z1 022430 0054 Rev. 00

Model(s): **Netti II**

## Parameters:

total length incl. footrests:	1100 mm
total length without footrests:	900 mm
width:	590 - 740 mm
height:	1120 mm
seat width:	350, 380, 400, 430, 450 and 500 mm
seat depth:	400, 425, 450, 475 and 500 mm
seat height:	460 and 490 mm (w/o cushion)
backrest height:	480 mm
seat plane angle:	-9° - +16° (manually adjustable)
backrest angle:	83°, 97°, 107° (fixed)
max. user weight:	125 kg
brake system:	knee lever brake (user brake)
back wheels: - size - tyres  - adjustable - fixation	24 x 1" Solid PU (puncture proof) Yes Quick release
front wheels: - size - tyres  - adjustable - fixation	Ø 175 x 33 mm Solid PU (puncture proof) Yes Quick release
frame material:	aluminum, steel
armrest:	height adjustable and removable
legrest:	height adjustable, angle adjustable and removable
anti tip device	adjustable

Tested according to: EN 12183:2014

Production Facility(ies): 057370

TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD TÜV SÜD  
 ZERTIFIKAT ◆ CERTIFICATE ◆ 認證證書 ◆ CERTIFICADO ◆ CERTIFICAT

A4 / 07.17