

Alu Rehab AS
Bedriftsvegen 23
4353 Klepp Stasjon
Norway

T: +47 51 78 62 20
post@My-Netti.com

Org.nr: NO 941 181 953 MVA
Bank: 9688 05 00047

TEST DECLARATION

March 2024

Netti AdaptPro manual wheelchair

- is tested and approved according to the European manual wheelchair test standard EN 12183 by accredited test laboratory. Max user weight **135 kg**.
Test report is available.
- is tested according to the international crash test standard for wheelchairs - ISO 7176-19 and is approved for being used as a seat in a car.
Max user weight when used as seat in car is 136 kg.
- is manufactured according to the international quality management system QMS standard for medical devices - ISO 13485. The quality system is yearly revised by accredited Notify Body. The certificate is available on our home page www.my-netti.com
- is CE marked and followed by an EU DoC - Declaration of Conformity - according to MDR 2017/245. The DoC is available on <https://www.my-netti.com>