

To whom it may concern

www.my-netti.com

post-klepp@meyragroup.com

T: +47 51 78 62 20

Alu Rehab AS
Bedriftsveien 23
N-4353 Klepp stasjon

February 2025

TEST DECLARATION

Netti Dynamic Base manual wheelchair

- is tested and approved according to the European manual wheelchair test standard EN 12183 by accredited test laboratory. Max user weight **135 kg**.
Test report is available.
- is tested according to the international crash test standard for wheelchairs - ISO 7176-19 and is approved for being used as a seat in a car.
Max user weight when used as seat in car is 135 kg.
- is manufactured according to the international quality management system QMS standard for medical devices - ISO 13485. The quality system is yearly revised by accredited Notify Body. The certificate is available on our home page www.my-netti.com.
- is CE marked and followed by an EU DoC - Declaration of Conformity - according to MDR 2017/245. The DoC is available on <https://www.my-netti.com>.