

To whom it may concern

[www.my-netti.com](http://www.my-netti.com)

[post-klepp@meyragroup.com](mailto:post-klepp@meyragroup.com)

T: +47 51 78 62 20

Alu Rehab AS  
Bedriftsveien 23  
N-4353 Klepp stasjon

February 2025

## TEST DECLARATION

### Netti V EL manual wheelchair

- is tested and approved according to the European manual wheelchair test standard EN 12184 by accredited test laboratory. Max user weight **135 kg**. Test report is available.
- is tested according to the international crash test standard for wheelchairs - ISO 7176-19 and is approved for being used as a seat in a car. Max user weight when used as seat in car is 135 kg.
- is manufactured according to the international quality management system QMS standard for medical devices - ISO 13485. The quality system is yearly revised by accredited Notify Body. The certificate is available on our home page [www.my-netti.com](http://www.my-netti.com).
- is CE marked and followed by an EU DoC - Declaration of Conformity - according to MDR 2017/245. The DoC is available on <https://www.my-netti.com>.