

# Product Combination agreement

## Alu Rehab A.S

Bedriftsvegen 23  
4353 Klepp Stasjon  
NORWAY

## Aspen Seating LLC

Ride Design  
8100 South Park Way, C400  
Littleton, CO 80120  
USA

as manufacturers of:

Netti manual Wheelchairs:

Netti II family, Netti III family  
Netti 4U family, Netti Dynamic System  
Netti Dynamic AdaptPro, Netti AdaptPro

Wheelchair cushions and backs:

Forward seat cushion,  
Java seat cushions  
Java Back, Decaf Back

hereby declare that the above-mentioned products can be combined under the following conditions:

1) The Velcro fixation on the cushions must fit with Velcro fixation on the wheelchairs.

*The Velcro system is meant to fasten the cushions to the wheelchair to secure the user of the wheelchair from sliding and to keep a stable posture position.*

2) It is the responsibility of the person fitting the wheelchair to the user to secure that the pivot point of the wheelchair is not moved in a critical position.

*As the cushions vary in thickness, design, height, width and length, it may change the properties of the wheelchair from its original specifications and risk assessment for the use of the wheelchair which needs to be considered by the caretaker doing the fitting.*

3) The user instruction for each product must be read carefully.


*The user instruction contains important information and warnings to make the products to be safe to use for the user.*

1. This agreement between *Alu Rehab AS* and *Aspen Seating LLC* (the Parties) states that the combination of the *Ride Design* Postural Support Devices and *Alu Rehab AS* Netti wheelchairs in combination (Products) meets the Medical Device Directive Council Directive 93/42/EEC and MDR Medical Device Regulation EU 2017/745 concerning product combination.
2. Aspen seating has performed the required risk evaluation of the Ride design cushions and Alu Rehab has performed the risk evaluation for Netti wheelchairs according to the MDD and MDR (Medical Device directory EEC 93/42/EEC and Medical Device Regulation EU2017/745). Both parties confirm that there are no increased risks by the combination of the Ride Design cushions and Netti wheelchairs.
3. The Products are both separately CE-marked devices and registered separately at a notified body.
4. Risk evaluation of the combination of Ride Design cushions and other wheelchairs must be performed by the executer of the combination.

5. *Alu Rehab AS* is responsible for the risk analysis of the combination of the medical device(s) listed above. The parties are responsible for possible limitations in use, instructions for installation and use, and required documentation for the medical device(s) listed above. *Alu Rehab AS* pledges to inform *Aspen Seating LLC* about any design changes to its wheelchairs which may affect the safety and or function of the Products in combination. *Aspen Seating LLC* pledges to inform *Alu Rehab AS* about any design changes to its postural support devices which may affect the safety and or function of the Products in combination.
6. Each Party has the product liability for its medical device. The Parties jointly have the product liability for the Products in combination.
7. The Parties renounce responsibility for a combination if not assembled according to each Parties installation and use instructions.
8. This agreement is valid from today and is subject to three months' notice for termination.
9. This document has been drawn up in two copies of which each Party has taken one.

The medical devices listed have been tested in combination in accordance with the Medical Device Directive Council Directive 93/42/EEC and Medical Device Regulation 2017/745.

**Alu Rehab AS**

Signature 

Title CEO

Place / Date Klepp Stasjon - 2021-02-18



**Alu Rehab AS**  
Bedriftsveien 23  
NO-4353 Klepp stasjon  
Norway  
NO 941 181 953 MVA

**Aspen Seating LLC**

Signature 

Title CEO

Place / Date LITTLETON, CO

2021-02-18